Credence Cartridge Radiomics Phantom CT Scans (CC-Radiomics-Phantom)

Redirection Notice

This page will redirect to https://www.cancerimagingarchive.net/collection/cc-radiomics-phantom/ in about 5 seconds.

Summary

This collection consists of 17 CT scans of the Credence Cartridge Radiomics (CCR) phantom, which was designed for use in studies of texture feature robustness. The scans were acquired at four medical centers using each center's chest protocol and were taken using GE (7 scans), Philips (5 scans), Siemens (2 scans), and Toshiba (3 scans) scanners. The CCR phantom has 10 cartridges, each with a unique texture, Fig 1. The first four cartridges are 3D printed ABS plastic with 20%, 30%, 40%, and 50% honeycomb fill, and they provide regular, periodic textures. The next three cartridges provide natural textures: sycamore wood, cork, and extra dense cork. A cartridges of shredded rubber particles provides textures similar to those of non-small cell lung cancer. The ninth cartridge is solid, homogenous acrylic and provides a minimal texture control. Finally, the 10th cartridge is 3D printed plaster has the highest electron density (400 – 600 HU) and is intended to more similar to bone.

In addition to the DICOM images for the 17 scans, this collection also contains two sets of contours as DICOM RT structure files. The first set provides $8x8x2 \text{ cm}^3$ contours for each cartridge in each scan. The second set provides 16 adjacent 2x2x2 cm³ contours for each cartridge in each scan. Researchers studying radiomics will be able to evaluate features for robustness across a variety of scanners. Features can be calculated using the researchers own software or third party software such as IBEX (imaging biomarker explorer).

Related publications: http://journals.lww.com/investigativeradiology/Abstract/2015 /11000/Measuring_Computed_Tomography_Scanner_Variability.3.aspx

The following paper was generated on different imaging modalities but the same phantom, this is a related but independent paper with a different set of authors: https://doi.org/10.1118/1.4934826

20% Filled ABS

30% Filled ABS

40% Filled ABS

50% Filled ABS

Sycamore Wood

Rubber Particles

Dense Cork

Solid Acrylic



Fig. 1: The 10 cartridges of the CCR radiomics phantom.

Acknowledgments

This data set was provided to TCIA by Authors: Mackin, Dennis; Fave, Xenia; Zhang, Lifei; Fried, David; Yang, Jinzhong; Taylor, Brian; Rodriguez-Rivera, Edgardo; Dodge, Cristina; Jones, Aaron Kyle; and Court, Laurence.

Data Access Data Access

Data Type	Download all or Query/Filter	License
Images and Radiation Therapy Structures (DICOM, 1.33 GB)	Download Search	CC BY 3.0
	(Download requires the NBIA Data	
	Keulevel)	

Click the Versions tab for more info about data releases.

Detailed Description Detailed Description

Collection Statistics	Radiology Image Statistics					
Modalities	CT, RTSTRUCT					
Number of Patients	17					
Number of Studies	17					
Number of Series	51					
Number of Images	2672					

1.33

Supporting Documentation and Metadata

Acquisition parameters for the phantom scans in this Collection:

Scan	Manufa cturer	Model	Kern el	Ty pe	Slice Thickness (mm)	Pixel (mm)	Spiral Pitch Factor	k Vp	Effectiv e mAs	CTDIvol (mGy)
CCR1 -GE1	GE	Discovery CT750 HD	stand ard	hel ical	2.5	0.49	0.98	1 20	81	6.19
CCR1 -GE2	GE	Discovery CT750 HD	stand ard	axi al	2.5	0.70	1.00	1 20	300	
CCR1 -GE3	GE	Discovery CT750 HD	stand ard	hel ical	2.5	0.78	0.98	1 20	122	9.3
CCR1 -GE4	GE	Discovery ST	stand ard	hel ical	2.5	0.98	1.35	1 20	143	16.3
CCR1 -GE5	GE	LightSpeed RT	stand ard	hel ical	2.5	0.98	0.75	1 20	1102	53.6
CCR1 -GE6	GE	LightSpeed RT16	stand ard	hel ical	2.5	0.98	0.94	1 20	367	18.8
CCR1 -GE7	GE	LightSpeed VCT	stand ard	hel ical	2.5	0.74	0.98	1 20	82	
CCR1 -P1	Philips	Brilliance Big Bore	В	hel ical	3.0	0.98	0.94	1 20	320	17.8
CCR1 -P2	Philips	Brilliance Big Bore	С	hel ical	3.0	0.98	0.94	1 20	369	15.8
CCR1 -P3	Philips	Brilliance Big Bore	В	hel ical	3.0	1.04	0.81	1 20	320	19.9
CCR1 -P4	Philips	Brilliance Big Bore	В	hel ical	3.0	1.04	0.81	1 20	369	19.9
CCR1 -P5	Philips	Brilliance 64	В	hel ical	3.0	0.98	0.67	1 20	372	16.4
CCR1 -S1	Siemens	Sensation Open	B31s	axi al	3.0	0.54	1.00	1 20	26 - 70	1.5
CCR1 -S2	Siemens	SOMATOM Definition Flash	['I70f ', '2']	hel ical	2.0	0.52	0.60	1 20	17 - 28	
CCR1 -T1	Toshiba	Aquilion	FC18	hel ical	3.0	0.63	1.11	1 20	135	4.0
CCR1 -T2	Toshiba	Aquilion	FC18	hel ical	3.0	0.63	1.11	1 20	135	3.8

CCR1	Toshiba	Aquilion ONE	FC18	hel	3.0	0.98	0.99	1	151	13.5
-T3				ical				20		

Citations & Data Usage Policy

Citations & Data Usage Policy

Users must abide by the TCIA Data Usage Policy and Restrictions. Attribution should include references to the following citations:

Data Citation

Mackin, D., Ray, X., Zhang, L., Fried, D., Yang, J., Taylor, B., Rodriguez-Rivera, E., Dodge, C., Jones, A., & Court, L. (2017). **Data From Credence Cartridge Radiomics Phantom CT Scans (CC-Radiomics-Phantom) [Data set].** The Cancer Imaging Archive. https://doi.org/10.7937/K9/TCIA.2017.zuzrml5b

① TCIA Citation

Clark, K., Vendt, B., Smith, K., Freymann, J., Kirby, J., Koppel, P., Moore, S., Phillips, S., Maffitt, D., Pringle, M., Tarbox, L., & Prior, F. (2013). The Cancer Imaging Archive (TCIA): Maintaining and Operating a **Public Information Repository.** In Journal of Digital Imaging (Vol. 26, Issue 6, pp. 1045–1057). Springer Science and Business Media LLC. https://doi.org/10.1007/s10278-013-9622-7 PMCID: PMC3824915

Additional Publication Resources:

The Collection authors suggest the below will give context to this dataset:

- Mackin, D., Fave, X., Zhang, L., Fried, D., Yang, J., Taylor, B., Rodriguez-Rivera, E., Dodge, C., Jones, A. K., & Court, L. (2015). Measuring Computed Tomography Scanner Variability of Radiomics Features. In Investigative Radiology (Vol. 50, Issue 11, pp. 757–765). Ovid Technologies (Wolters Kluwer Health). https://doi .org/10.1097/rli.000000000000180
- The following paper was generated on different imaging modalities but the same phantom, this is a related but independent paper with a different set of authors: Fave, X., Mackin, D., Yang, J., Zhang, J., Fried, D., Balter, P., Followill, D., Gomez, D., Kyle Jones, A., Stingo, F., Fontenot, J., & Court, L. (2015). Can radiomics features be reproducibly measured from CBCT images for patients with non-small cell lung cancer? In Medical Physics (Vol. 42, Issue 12, pp. 6784–6797). Wiley. htt ps://doi.org/10.1118/1.4934826

Other Publications Using This Data

TCIA maintains a list of publications which leverage our data. If you have a manuscript you'd like to add please contact TCIA's Helpdesk.

Versions

Version 1 (Current): Updated 2017/07/28

Data Type	Download all or Query/Filter	License
Images (DICOM, 1.33 GB)	Download Search (Requires NBIA Data Retriever.)	CC BY 3.0